



Medica Corporation
 5 Oak Park Drive
 Bedford, Massachusetts 01730
 Tel 781 275 4892
 Fax 781 275 2731
 www.mediacorp.com

Products For Health Care

CERTIFICATE OF COMPLIANCE/ANALYSIS

This is to certify that the material shipped on the order is in conformance with MEDICA specifications for this product manufactured on 05/22/2019 with the following analysis:

Product Description: Calcium Reference Electrode

Medica Product ID: 2152

MSDS Available: No

Lot Number: 19142

Install Date: 2020-05-22

<u>Product Test</u>	<u>Units</u>	<u>Specification</u>	<u>Results</u>	<u>Test Procedure No.</u>
Slope Test	N/A	Pass Test	Pass	TP001958-001
Control Recovery	N/A	Pass Test	Pass	TP001958-001

QC or QA Signature: *Jeanne Jarvis*

Date: 06-02-19

Important to Purchaser:
 "The following is made in lieu of all expressed warranties: Seller's only obligation shall be to replace such quantity of the product proved to be defective or to refund the purchase price. Seller shall not be liable for the injury, loss or damage, direct or consequential, arising out of the use or inability to use the product. Before using, user shall determine the suitability of the product for the intended use and user assumes all risk and liability whatsoever in connection therewith. The foregoing may not be changed except by an agreement by an officer of seller."