



Medica Corporation  
 5 Oak Park Drive  
 Bedford, Massachusetts 01730  
 Tel 781 275 4892  
 Fax 781 275 2731  
 www.medicacorp.com

Products For Health Care

**CERTIFICATE OF COMPLIANCE/ANALYSIS**

This is to certify that the material shipped on the order is in conformance with MEDICA specifications for this product manufactured on 11/16/2018 with the following analysis:

**Medica Product ID: 6201**

**Product Description: EBG/ES pH Sensor**

**MSDS Available: No**

**Lot Number: 18111047**

**Install Date: 2019-11-16**

<u>Product Test</u>	<u>Units</u>	<u>Specification</u>	<u>Results</u>	<u>Test Procedure No.</u>
Slope Test	mV/Decade	59 to 64.5	Pass	TP004302-001
Millivolt Test	millivolts	-100 to +100	Pass	TP004302-001
Control Test	N/A	Pass Test	Pass	TP004302-001

**QC or QA Signature:**  **Date:** 12-07-18

Important to Purchaser:  
 "The following is made in lieu of all expressed warranties: Seller's only obligation shall be to replace such quantity of the product proved to be defective or to refund the purchase price. Seller shall not be liable for the injury, loss or damage, direct or consequential, arising out of the use or inability to use the product. Before using, user shall determine the suitability of the product for the intended use and user assumes all risk and liability whatsoever in connection therewith. The foregoing may not be changed except by an agreement by an officer of seller."